TATA MEMORIAL CENTRE Advanced Centre for Treatment, Research and Education in Cancer

Technical Specification of: LED X-RAY View box

Name of the Vendor:

X-Ray view box, which should be capable for viewing 3 x-rays

TECHNICAL REQUIREMENT	<u> </u>
Name of the Manufacturer:_	

Sr. No.	Description of item	Compliance
	Model Name:	
	Make:	
1	X-Ray view box should be LED illumination type.	
2	View box should be capable for viewing 3 nos. of	
	17X14 inch X-ray films side by side.	
3	Total viewing area should be 110X 44 cm or	
	more.	
4	Total frame size should be approx. 125X 60 cm	
5	Thickness of the view box should be <5cm	
6	Power supply should be inbuilt.	
7	Brightness control & indication should be	
	available.	
8	Maximum intensity should be > 10000 lux	
9	Input power supply should be 220-240V/ 50Hz,	
	Indian type plug should be provided.	
10	ON/OFF switch should be available on the view	
	box.	
11	Power consumption should be <50W	
12	Film holder should be available	
13.	Vendor responsibility to fix it in our OTs	
	Warranty: 1 year	
	Free Service: NA	