

Department- ,	Pathology	
Technical Specifications- Cabinet Block Filing		
Sr no	Item Description	

Please quote the model which is including the latest state of art equipment and meets the basic minimum specifications. The quoted model should be latest introduced with end of life not before 12 years from the date of installation at TMC.

Model Name:
Make:
Year of Introduction:
Probable end of support:

Sn	Item Description
1	Should have convenient storage and retrieval of paraffin blocks
2	Should be constructed from good quality cold rolled steel with powder coating finish.
3	Should be stackable
4	Each drawer should have a back stop
5	Each cabinet should have 7 drawers and each drawer should have 14 columns accommodating 50 blocks in each column . Each drawer should have capacity of 700 blocks , likewise each cabinet should have 7 drawers accommodating 4900 blocks.
6	Size of each cabinet should be W 20.5" X D 19" X H 19"
7	Size of each drawer should be W 18.5" X D 18" X H 2.25"
8	Width of each column should be 1.25"
9	Should have writable socket/stripes on drawers for labeling
10	Drawers of cabinet should be lockable for protection and security with 1 master key for all cabinets
11	Specify pre installation requirements (electrical, HVAC etc)
12	Specify its weight.
13	space
14	tender
15	AMC(includes unlimited breakdown calls)
16	Demo of the quoted model , will be required if so desired by the user, after the opening of the technical bid and prior to opening of financial bid. This is for technical evaluation.
17	engineer for attending the breakdown calls.
18	complaint on toll free or by email.
1	Warranty- 2 years
2	after warranty
3	minimum 10 years and to be submitted to HBCHRC, VIZAG