

TATA MEMORIAL CENTRE			
ADVANCED CENTRE FOR TREATMENT, RESEARCH & EDUCATION IN CANCER			
TECHNICAL SPECIFICATION OF :-Single Channel Variable Volume Pipettes			
A)	Name of the Vendor :		
SR. NO.	Technical Specifications / Scope of supply	Quantity	<b>COMPLIANCE</b> 1) Mention "Complied /Not Complied" 2) Highlight any deviations. 3) Mention part number/Catalogue number of the relevant item quoted.
B)	Name of the Manufacturer :		
C)	Name of the Model:		
D)	Make:		
E)	Model No. of the Equipment (If the equipment is a combination of more than one equipment then list down the model no. pertaining to each) Can Quote more than one model.:		
F)	Year of Introduction Internationally		
G)	Year of Introduction in India		
H)	Technical Specifications/Scope of supply		
I)	Quote for Latest state of the art model and other models , if any, meeting the tender specifications.		
1	Single Channel Variable Volume Pipettes, complete as per below mentioned configuration and specifications:	10 Nos	
2	Ergonomic design, rugged and sturdy pipette. Suitable for multi-user environment.		
3	Air displacement type, single channel variable volume pipettes.		
4	Continuously variable volume setting knob. Smooth operation without the cracking sound during volume setting.		
5	Volume setting knob should be at the top.		
6	Metallic plunger, piston and tip ejector.		
7	Volume locking mechanism should be available.		
8	Adjustable tip ejector pressing position.		
9	Fully autoclavable.		
10	Qty. required for single Channel Pipettes as per following:		
11	0.2 to 2 microliter- Max. permissible random error 0.04 microliter or less as per ISO 8655	2 Nos	
12	1.0 to 10 microliter-Max. permissible random error 0.08 microliter or less as per ISO 8655	2 Nos	
13	2.0 to 20 microliter-Max. permissible random error 0.10 microliter or less as per ISO 8655	2 Nos	
14	20 to 100 microliter-Max. permissible random error 0.60 microliter or less as per ISO 8655	2 Nos	
15	100 to 1000 microliter-Max. permissible random error 3.0 microliter or less as per ISO 8655	2 Nos	
16	Other terms:		
17	Relevant literature/documents should be provided with the technical bid in respect to the random error values.		Attached/Not attached
18	Product literature, Brochure should be provided with the technical bid.		Attached/Not attached
19	Warranty: 2 years comprehensive-on site- Factory warranty-		Agree/Not Agree
20	Important terms to be noted by the bidders:		
21	Declaration by the bidder		
22	We have quoted for all the items meeting the description/scope of supply in the Financial bid as per prescribed format of the Tender documents and we agree that Partial/incomplete offers are liable for rejection.		Agree/Not Agree