

NAME OF THE EQUIPMENT: Patient Warmer**TECHNICAL REQUIREMENT:**

Warming system for Patient body warming through disposable mattresses.

Can Quote more than one model_

Name of the Manufacturer: _____

Name of the Vendor: _____

Sr. No.	SPECIFICATIONS	VENDOR'S OFFER (all rows to be filled in detail)
A	Name of the Model:	
B	Model No. of the Equipment (If the equipment is a combination of more than one equipment then list down the model no. pertaining to each) :	
C	Year of Introduction Internationally	
D	Year of Introduction in India	
E	Country of Origin	
F	Specification	
1	Temperature range from 36-44deg C or better	
2	Display of temperature	
3	Over temperature alarm should be there.	
4	Should be IV pole mount and floor mount	
5	Air filter for clean air should be available.	
6	Automatic over temperature shut off facility should be available	
7	Electrical connection should be 220V-240V, 50Hz	
8	Should be match with electrical safety standards	
9	Convection heating is preferable.	
10	FDA and CE approved	
11	Standard accessories (Which comes along with the machine, pricing of which is part of the equipment) Please list down the Standard accessories	
12	Tubing/Hose- 1 no. each	
13	Mattress- Atleast 2 nos.	
14	Essential Accessories (Accessories which are required for the equipments functional needs but is not part of the standard equipment). Please list down the Essential accessories and with prices in the Financial bid.	
15	Optional accessories (List out all the optional attachments & accessories compatible with the specified model and with prices in the Financial bid)	
16	Any other salient feature you want to convey about the equipment	
17	Essential spare parts & consumables: Provide List of required spares and consumables (if any) with their list price and % discount offered in the Financial bid	
18	Mattress price should be provided in price bid. And the price may be freezed for a period of two years.	
19	Installation:	

	Company trained personnel as per end user's satisfaction should do installation and impart training to end user.	
20	Training of staff: Provide necessary training to Institute's staff / engineers in routine maintenance and fault finding.	
21	Service support: Complete operational manuals should be provided. Company should provide technical support for hardware and software, spares and consumables. Company should provide required spares and consumables for at least 7 years even after discontinuation of the model.	
22	Warranty: at least 2 years warranty is a mandatory requirement. Should have local service facility	
23	Free Annual Maintenance Contract is mandatory for 5 years post warranty	
24	Safety requirements: It should follow International safety requirement	
25	User's list: A list of installations with the address and contact numbers to be provided. (User list should be for the quoted model)	
26	Multiple Models: If more than one model is offered, then they are to be separately filled and submitted according to the given format.	