

Tata Memorial Centre			
Advanced Centre for Treatment, Research and Education in Cancer			
Technical Specification of : Petriplate Warmer			
Sr.no.	Description of Item	Quantity	COMPLIANCE 1) Compliance column may be filled with relevant data, figures, range etc. as applicable . Do not just mention "YES / NO / Complied". 2) Highlight any deviations . 3) Mention part
A)	Name of the Vendor :		
B)	Name of the Manufacturer :		
C)	Name of the Model:		
D)	Year of Introduction Internationally		
E)	Year of Introduction in India		
	Technical Specifications / Scope of Supply		
1	Petriplate Warmer-complete as per below mentioned configuration and specifications:-	1 No.	
2	Equipment should be compact, with small footprint and rugged in construction.		
3	Suitable for warming culture plates, petri dishes and reagents.		
4	Digital display for temperature.Keypad for user settings.		
5	Temperature setting: Ambient+5 to 45 deg. cel. Or better		

6	Anodized aluminium surface. Usable dimentions: 7 inch x 8 inches or more.		
7	Single unit. Controller and heating plate should be housed in a single housing. (DO not quote models with plate separate and controller separate)		
8	Countdown timer: 15 min. or more.		
9	Input power supply requirements: 240 V AC ± 10 %, 50 Hz, single phase		
10	Warranty: 2 years comprehensive-on site-Factory warranty- for entire equipment.		Agree/Not Agree
11	After Sales Service Support.		
12	Manufacturer should have its authorised service provider's office located in Mumbai, Navi Mumbai,Thane Area. Service provider should have adequare strength of service personnel trained on servicing and maintenance of the quoted model. Bidders who do not qualify this critera shall not be considered. Bidder should provide address of the authorized service centre and details of number of personnel with training credentials.		
13	Important terms to be noted by the bidders:		
14	<u>Read the above scope of supply carefully and quote accordingly. Incomplete and /or partially complete offers are liable to be rejected.</u>		Agree/Not Agree
15	Item/Equipment quoted should not be a second hand/refurbished/on verge of obsolescence/outdated model. Similarly, after award of contract, the item supplied should be a newly manufactured item and not the old item already manufactured and stocked for more than 1 month.		Agree/Not Agree

16	<p>After opening of the Technical bid (Part-1), Physical demonstration of the quoted model may have to be shown / arranged by the bidder, if requested by the Institute. Physical demonstration should be arranged at ACTREC within 15 days from the date of request. Physical Demonstration may be requested to confirm the availability of any or all technical features as mentioned/stated in the technical bid. Physical Demonstration will also be a part of technical evaluation process. If the bidder does not comply, such bids are liable to be disqualified. (Demonstration of quoted model is to be shown and not the demonstration of similar models with different technical specifications and features)</p>		Agree/Not Agree
17	Past experience of the bidders in terms of quality of supply		Agree/Not Agree
18	<p>Complete and detailed information should be provided in respect to each point specified in the specifications. <u>Technical bids that are incomplete in any respect are liable to be rejected.</u> Provide relevant supportive information, publications, catalogue, etc. Bidders providing misleading or wrong information are liable to be rejected. All technical claims should be printed in the technical brochure of the equipment.</p>		Agree/Not Agree
19	<p>If any contradictory statements /figures/information is observed in the compliance chart and in the technical bid, then the technical information mentioned in the product literature/brochure will be considered true and further evaluation will be done based on the information given in the product literature/brochure. Regarding the warranty and free AMC terms, information quoted in the technical bid will only be taken as valid terms.</p>		Agree/Not Agree
20	<p>Compliance column may be filled with relevant data, figures, range etc. as applicable. Do not just mention "YES / NO / Complied".</p>		Agree/Not Agree

21	Provide list of current installations(in the last 2-3 years) of the quoted model in India. Contact details of the end users should be provided. Provide name of the end user, email ID, tel. no. etc.		Agree/Not Agree
22	Declaration by the bidder		
23	We have quoted for all the items meeting the description/scope of supply in the Financial bid as per prescribed format of the Tender documents and we agree that Partial/incomplete offers are liable for rejection.		Yes/No