

	Tata Memorial Centre		
	Advanced Centre for Treatment, Research and Education in Cancer		
	Technical Specifications Of: Heating water bath		
Sr No	Scope of supply	Quantity	COMPLIANCE 1) Compliance column may be filled with relevant data, figures, range etc. as applicable. Do not just mention "YES / NO / Complied". 2) Highlight any deviations. 3) Mention part number/Catalogue number of the relevant item quoted.
A)	Name of the Vendor :		
B)	Name of the Manufacturer :		
C)	Name of the Model:		
D)	Year of Introduction Internationally		
E)	Year of Introduction in India		
	Technical Specifications / Scope of Supply		
1	Heating Water Bath , complete as per below mentioned configuration and specifications:	1 No	
2	Compact, tabletop model.		
3	General Purpose water bath. Epoxy powder coated exterior and stainless steel interior.		
4	Effective, usable Bath capacity: 24-26 Ltr Rectangular bath chamber. (Do not quote models more than 26 Ltr. capacity)		
5	Seamless stainless steel tank, corrosion resistant.		
6	Bath chamber should be rectangular in shape and should be suitable to keep test tube racks, bottles etc. directly on the bottom of the chamber. Hence Heaters should not be exposed. It should be covered with diffuser tray.		
7	Temperature range: Ambient+ 5 deg. cel to +99 deg. cel. or more		
8	PID Temperature controller. Digital display for set and run parameters. Precise temperature control. (provide confirmatory document)		
9	Temperature accuracy error: +/- 0.2 deg. Cel. or lower		
10	The bath should be supplied with dome shaped lid. Lid should be hinged to the bath. (provide real photograph of the lid with the bath.)		
11	Circulator not needed. DO not quote models with stirrer/circulator		
12	Heating capacity: 1000 watts or more		
13	Safety features needed:		
14	Over temperature alarm.		
15	Additional Thermostat for safety ageing over temperature. (provide confirmatory document)		
16	Attach product literature/brochure with the technical bid.		
17	Input power supply requirements: 240 V AC ± 10 %, 50 Hz, single phase		
18	Warranty: 2 years comprehensive-on site- Factory warranty- for entire equipment.		
19	After Sales Service:		
20	Manufacturer or its Authorised service provider should have its office in Mumbai/Navi Mumbai/Thane Area. Provide details of the office address, number of technical qualified staff available at local office.		

21	Important terms to be noted by the bidders:		
22	<u>Read the above scope of supply carefully and quote accordingly.</u> <u>Incomplete and /or partially complete offers are liable to be rejected.</u>		Agree/Not Agree
23	Item/Equipment quoted should not be a second hand/refurbished/on verge of obsolescence/outdated model. Similarly, after award of contract, the item supplied should be a newly manufactured item and not the old item already manufactured and stocked for more than 1 month.		Agree/Not Agree
24	After opening of the Technical bid (Part-1), Physical demonstration of the quoted model may have to be shown / arranged by the bidder, if requested by the Institute. Physical demonstration may be shown at ACTREC within 15 days from the date of request. Physical Demonstration may be requested to confirm the availability of any or all technical features as mentioned/stated in the technical bid. Physical Demonstration will also be a part of technical evaluation process. If the bidder does not comply, such bids are liable to be disqualified. (Demonstration of quoted model is to be shown and not the demonstration of similar models with different technical specifications and features)		Agree/Not Agree
25	Past experience of the bidders in terms of quality of supplied equipments, after sales service and application support will be taken into consideration while technical evaluation.. Bidders who has unsatisfactory past experience in last 2-3 years, in terms of quality of supplied equipments, after sales service and application support, bids of such bidders may liable to be rejected.		Agree/Not Agree
26	Unpacking and Shifting the consignment to the installation site is to be included in the scope of supply. Bidder should take responsibility to lift/shift the consignment from unloading site to the installation site . If needed, Bidder has to arrange for the labourers at no charge to ACTREC. (Bidders may visit ACTREC to know unloading site and installation site)		Agree/Not Agree
27	Complete and detailed information should be provided in respect to each point specified in the specifications. <u>Technical bids that are incomplete in any respect are liable to be rejected.</u> Provide relevant supportive information, publications, catalogue, etc. Bidders providing misleading or wrong information are liable to be rejected. All technical claims should be printed in the technical brochure of the equipment.		Agree/Not Agree
28	If any contradictory statements /figures/information is observed in the compliance chart and in the technical bid, then the technical information mentioned in the product literature/brochure will be considered true and further evaluation will be done based on the information given in the product literature/brochure. Regarding the warranty and free AMC terms, information quoted in the technical bid will only be taken as valid terms.		Agree/Not Agree
29	Compliance column may be filled with relevant data, figures, range etc. as applicable. Do not just mention "YES / NO / Complied " .		Agree/Not Agree
30	Provide list of current installations(in the last 2-3 years) of the quoted model in India. Contact details of the end users should be provided. Provide name of the end user, email ID, tel. no. etc.		Agree/Not Agree
31	Declaration by the bidder		
32	We have quoted for all the items meeting the description/scope of supply in the Financial bid as per prescribed format of the Tender documents and we agree that Partial/incomplete offers are liable for rejection.		Yes/No