

Tata Memorial Centre			
Advanced Centre for Treatment, Research and Education in Cancer			
Technical Specification of : Laser Particle Counter			
Sr.no.	Description of Item	Quantity	<b>COMPLIANCE</b> 1) Compliance column may be filled with relevant data, figures, range etc. as applicable. Do not just mention "YES / NO / Complied ". 2) <b>Highlight any deviations.</b> 3) <b>Mention part number/Catalogue number of the relevant item quoted.</b>
	<b>Technical Specifications / Scope of Supply</b>		
1	<b>Laser Particle Counter, Complete as per below mentioned configuration and specifications:-</b>	<b>1 No.</b>	
2	Handheld monitor suitable for air particle counter.		
3	Suitable for measuring 0.3 micron to 5 micron particles or more.		
4	3 channels or more.		
5	Backlit LCD display for reading.		
6	Battery operated.		
7	<b>Warranty:</b> 1 year- standard manufacturer's warranty		Agree/Not Agree
8	<b>Declaration by the bidder</b>		
9	We have quoted for all the items meeting the description/scope of supply in the Financial bid as per prescribed format of the Tender documents and we agree that Partial/incomplete offers are liable for rejection.		Yes/No