

NAME OF THE EQUIPMENT: 30 Degree Laproscope

TECHNICAL REQUIREMENT:

Intended use: 30 Degree scope need for Laparoscopic surgery

Can quote for more than one model

Name of the Manufacturer: _____

Name of the Vendor: _____

Sr. No.	SPECIFICATIONS	VENDOR'S OFFER (all rows to be filled in detail)
A	Name of the Model:	
B	Model No. of the Equipment (If the equipment is a combination of more than one equipment then list down the model no. pertaining to each) :	
C	Year of Introduction Internationally	
D	Year of Introduction in India	
E	Specification	
1	Autoclavable ultra high definition laparoscope	
2	Diameter should be 10mm,	
3	working length-325mm	
4	Angle 30 degree ,	
5	Field of view 100-108degree	
6	Warranty: Atleast 6 months against manufacturing defect	