

Tata Memorial Centre			
Advanced Centre for Treatment, Research and Education in Cancer			
Technical Specification of : Magnetic Stirrer 2 Nos.			
Sr.no.	Description of Item	Quantity	COMPLIANCE Mention "Complied /Not Complied" 2) Highlight any deviations. 3) Mention part number/Catalogue number of the relevant item quoted.
Technical Specifications / Scope of Supply			
1	<b>Magnetic Stirrers, complete as per below mentioned configuration and specifications:-</b>	2 Nos.	
2	Magnetic stirrer cum hot plate model should be quoted.		
3	Equipment should be compact and rugged in construction		
4	It should have electronic control/potentiometer for speed regulation as well as temperature setting for hot plate.		
5	Heating element should be built-in below the platform.		
6	It should be able to perform stirring with good speed regulation.		
7	Platform size: platform- approx. 18 cm x 18 cm		
8	Required Stirring speed range: ~ 100 to 1000 RPM		
9	Max. temperature range: upto approx. +100 deg. cel. Calibrated knob/dial for user required temperature setting.		
10	Platform material required : Glass ceramic, Anti corrosive ( SS platform models shall not be considered)		
11	Stirring operation should not be influenced by thickness of the glass.		
12	<b>After Sales Service Support.</b>		
13	Manufacturer should have its authorised service provider's office located in Mumbai, Navi Mumbai, Thane Area. Service provider should have adequate strength of service personnel trained on servicing and maintenance of the quoted model. Bidders who do not qualify this criteria shall not be considered. Bidder should provide address of the authorized service centre and details of number of personnel with training credentials.		
14	<b>Input power supply requirements:</b> 240 V AC $\pm$ 10 %, 50 Hz, single phase		
15	<b>Warranty:</b> 2 years comprehensive-on site- Factory warranty- for entire equipment.		
16	<b>Important terms to be noted by the bidders:</b>		
17	<u>Read the above scope of supply carefully and quote accordingly.</u> <u>Incomplete and /or partially complete offers are liable to be rejected.</u>		
18	Item/Equipment quoted should not be a second hand/refurbished/on verge of obsolescence/outdated model. Similarly, after award of contract, the item supplied should be a newly manufactured item and not the old item already manufactured and stocked for more than 1 month.		
19	After opening of the Technical bid (Part-1), Physical demonstration of the quoted model may have to be shown / arranged by the bidder, if requested by the Institute. Physical demonstration should be arranged at ACTREC within 15 days from the date of request. Physical Demonstration may be requested to confirm the availability of any or all technical features as mentioned/stated in the technical bid. Physical Demonstration will also be a part of technical evaluation process. If the bidder does not comply, such bids are liable to be disqualified. (Demonstration of quoted model is to be shown and not the demonstration of similar models with different technical specifications and features)		
20	Past experience of the bidders in terms of quality of supplied equipments, after sales service and application support will be taken into consideration while technical evaluation.. Bidders who has unsatisfactory past experience in last 2-3 years, in terms of quality of supplied equipments, after sales service and application support, bids of such bidders may liable to be rejected.		
21	<b>Unpacking and Shifting the consignment to the installation site is to be included in the scope of supply. Bidder should take responsibility to lift/shift the consignment from unloading site to the installation site . If needed, Bidder has to arrange for the labourers at no charge to ACTREC. ( Bidders may visit ACTREC to know unloading site and installation site)</b>		
22	Complete and detailed information should be provided in respect to each point specified in the specifications. <u>Technical bids that are incomplete in any respect are liable to be rejected.</u> Provide relevant supportive information, publications, catalogue, etc. Bidders providing misleading or wrong information are liable to be rejected. All technical claims should be printed in the technical brochure of the equipment.		

23	If any contradictory statements /figures/information is observed in the compliance chart and in the technical bid, then the technical information mentioned in the product literature/brochure will be considered true and further evaluation will be done based on the information given in the product literature/brochure. Regarding the warranty and free AMC terms, information quoted in the technical bid will only be taken as valid terms.		
24	Compliance column may be filled with relevant data, figures, range etc. as applicable. Do not just mention "YES / NO / Complied ".		
25	Provide list of current installations( in the last 2-3 years) of the quoted model in India. Contact details of the end users should be provided. Provide name of the end user, email ID, tel. no. etc.		
26	<b>Declaration by the bidder</b>		
27	We have quoted for all the items meeting the description/scope of supply in the Financial bid as per prescribed format of the Tender documents and we agree that Partial/incomplete offers are liable for rejection.		