

NAME OF THE EQUIPMENT: Pacemaker-External

TECHNICAL REQUIREMENT:

temporary external pacemaker is intended to be used in conjunction with a cardiac pacing lead system for temporary atrial or ventricular pacing in a clinical environment by trained personnel

Can quote for more than one model

Name of the Manufacturer: _____

Name of the Vendor: _____

Sr. No.	SPECIFICATIONS	VENDOR'S OFFER (all rows to be filled in detail)
A	Name of the Model:	
B	Model No. of the Equipment (If the equipment is a combination of more than one equipment then list down the model no. pertaining to each) :	
C	Country of Origin:	
D	Year of Introduction Internationally	
E	Year of Introduction in India	
F	Specification	
1	Pacemaker should be capable for Clinical use	
2	The temporary pacemaker can be used where short-term demand (synchronous) or asynchronous pacing is indicated for therapeutic, prophylactic, or diagnostic purposes	
3	Basic pacing rate should be 30-200 ppm or better	
4	Rapid Atrial pacing rate should be approximately 80-800ppm	
5	Output amplitude should be around 0.1-25mA	
6	Constant current — adjust pacing output voltage in response to changes in lead impedance	
7	Pulse width-1.5ms	
8	Sensitivity-0.4 to 20mV	
9	Compatible with unipolar and bipolar electrode	
10	Equipment should be compatible with readily available alkaline batteries like AA or AAA.	
11	Required leads should be provided with the pacemaker	
12	Any Additional/ salient features which you want to highlight pertaining to the model quoted	
13	Warranty: at least 2 years warranty is a mandatory requirement	
14	Multiple Models: If more then one model is offered, then they are to be separately filled and submitted according to the given format.	