

TECHNICAL COMPLIANCE FORM		
Ref: Tender No:		
TATA MEMORIAL CENTRE		
ADVANCED CENTRE FOR TREATMENT, RESEARCH & EDUCATION IN CANCER		
TECHNICAL SPECIFICATION OF :- - "Fluid Warmer- 16 Nos."		
Intended Use: For Anesthesia use. To warm the liquid and blood before infusing to the Patient .		
Sr. No.	Name of the Vendor :	
A	Technical Specifications / Scope of supply	COMPLIANCE 1) Mention "Complied /Not Complied" 2) Highlight any deviations. 3) Mention part number/Catalogue number of the relevant item quoted.
B	Name of the Manufacturer :	
C	Country of Origin:	
D	Model Name / Model No. of the Equipment (If the equipment is a combination of more than one equipment then list down the model no. pertaining to each)	
E	Year of Introduction Internationally	
F	Year of Introduction in India	
G	Please quote the latest state of the art model which meets the tender specifications. The offer should be complete in all respect for the intended use.	
H	Technical Specifications/Scope of supply	
I	" Fluid Warmer-16 Nos. ", complete as per below mentioned configuration and specifications:	
J	Features and specifications required:	
1	Equipment should be capable of warming Blood , other blood products and injecting fluids.	
2	Equipment should be capable of mounting on IV stand.	
3	Warmer temperature should stabilize within 15 minutes.	
4	Uniform heating should be available throughout the liquid	
5	Flow rate should be 50 to 500ml/ hr or better	
6	Fluid should be warm till it reaches the patient body.	
7	Temperature range of circulating fluid should be 39 deg C +_2 degC	
8	.Audible alarm for High temperature should be available	
9	Temperature display of circulating fluid should be available	
10	Operating Manual should be provided	
11	USFDA /EuropeanCE/ICMED/BIS Approval	
12	It should be compatible with international Patient safety/ Electrical safety standards with proof	
13	Upgradability capability (List down possible upgrades for the quoted model)	

14	Regulatory Approvals, If any. Details with copies of approvals.	
15	Safety requirements: It should follow International / national safety requirement. Please specify certification with certifying agency and country with copies of certificate.	
16	User's list: A list of installations with the address and contact numbers to be provided. (User list should be for the quoted model.)	
17	Input power supply requirements: 240 V AC \pm 10 %, 50 Hz, single phase. Specify if any other power supply requirement is recommended.	
18	After Sales Service Support.	
19	<p>1. Complete operational manuals and technical information should be provided.</p> <p>2. Company should provide technical support for hardware and software, spares and consumables.</p> <p>3. Company should provide required spares and consumables for at least 7 years even after discontinuation of the model.</p> <p>4. Bidders who do not qualify this criteria shall not be considered.</p> <p>Bidder should provide address of the authorized service centre and details of number of personnel with training credentials.</p>	
20	Installation, Commissioning, testing and Training	
21	Unpacking and Shifting the consignment to the installation site is to be included in the scope of supply. Bidder/manufacturer/authorized service provider should take responsibility to lift/shift the consignment from unloading site to the installation site . Unloading site shall be "Stores Department, KS Building, ACTREC Campus". If needed, Bidder has to arrange for the labourers at no charge to ACTREC. (Before submitting the quotation, bidders may visit ACTREC to know unloading site and installation site)	
22	Installation,Commissioning and Training is included in the scope of supply. Bidder, Manufacturer and/or its authorized representative should undertake installation and commissioning of the equipment.	
23	Warranty and after sales support	
24	Warranty: 2 years	
25	Free AMC -labour only: Free AMC for 5 years post warranty	
26	During AMC service period, manufacturer/authorized service provider should provide preventive services as well as attend the breakdown calls as and when requested by the end user/Institute on " no charge basis"	
27	Important terms to be noted by the bidders:	
28	<p><u>Read the above scope of supply carefully and quote accordingly.</u></p> <p><u>Incomplete and /or partially complete offers are liable to be rejected.</u></p>	Agree/Not Agree
29	Manufacturer/bidder should clearly indicate typical life span / Life expectancy of the quoted equipment.	-- years

30	Bidders should clearly mention about compliance of the warranty and AMC/free service terms in the technical bid. Bids not fully complying to the warranty and AMC/free service terms are liable to be rejected.	Agree/Not Agree
31	Item/Equipment quoted should not be a second hand/refurbished/on verge of obsolescence/outdated model. Similarly, after award of contract, the item supplied should be a newly manufactured item and not the old item already manufactured and stocked for more than 1 month.	Agree/Not Agree
32	After opening of the Technical bid (Part-1), Physical demonstration of the quoted model may have to be shown / arranged by the bidder, if requested by the Institute. Physical demonstration may be shown at one of the end user's site/Principle company's application lab/manufacturing site located in Mumbai/Navi Mumbai/Thane cities. If there are no installations of the quoted model in Mumbai/Navi Mumbai/Thane cities, then the quoted model may have to be brought in at ACTREC for demo purpose within 10 days from the date of request . Physical Demonstration may be requested to confirm the availability of any or all technical features as mentioned/stated in the technical bid. Physical Demonstration will also be a part of technical evaluation process. If the bidder does not comply, such bids are liable to be disqualified. (Demonstration of quoted model is to be shown and not the demonstration of similar models with different technical specifications and features)	Agree/Not Agree
33	Past experience of the bidders in terms of quality of supplied equipments, after sales service and application support will be taken into consideration while technical evaluation.. Bidders who has unsatisfactory past experience in last 2-3 years, in terms of quality of supplied equipments, after sales service and application support, bids of such bidders may liable to be rejected.	Agree/Not Agree
34	Complete and detailed information should be provided in respect to each point specified in the specifications. <u>Technical bids that are incomplete in any respect are liable to be rejected.</u> Provide relevant supportive information, publications, catalogue, etc. Bidders providing misleading or wrong information are liable to be rejected. All technical claims should be printed in the technical brochure of the equipment.	Agree/Not Agree
35	If any contradictory statements /figures/information is observed in the compliance chart and in the technical bid, then the technical information mentioned in the product literature/brochure will be considered true and further evaluation will be done based on the information given in the product literature/brochure. Regarding the warranty and free AMC terms, information quoted in the technical bid will only be taken as valid terms.	Agree/Not Agree
36	Compliance column may be filled with relevant data, figures, range etc. as applicable. Do not just mention "YES / NO / Complied ".	Agree/Not Agree

37	Declaration by the bidder	
38	We have quoted for all the items meeting the description/scope of supply in the Financial bid as per prescribed format of the Tender documents and we agree that Partial/incomplete offers are liable for rejection.	Yes/No