

**TATA MEMORIAL CENTRE ADVANCED CENTRE FOR TREATMETN,
RESEARCH AND EDUCATION IN CANCER (ACTREC)
PURCHASE DEPARTMENT**

ACTREC/PUR/2022-2023/E-TENDER/15

12.07.2022

CORRIGENDUM No-5

**Reference to CPPP Tender Notice ACTREC/PUR/22-23/E-TENDER/15 for
Supply, Installation, Testing and Commissioning of MRI 3T Scan system along with
required Turnkey Works. Qty 1 No.**

MODIFICATIONS IN TECHNICAL SPECIFICAION –

| Sr. No | EXSITING TECHNICAL SPECIFICATION | REFINED TENDER SPECIFICATION TO BE READ AS |
|--|--|---|
| 2C SHIM SYSTEM | Second order/High order shimming should be offered as standard | Second order or High order shimming or Auto shimming should be offered as standard |
| 3A GRADIENT SYSTEM | Field of View should be at least 55/55/50cm in all three axis | Field of View should be at least 50/50/45 cm in all three axis or better |
| 8M83 - SUB POINT AK ADDITIONAL WORKSTATION | MRI Finger printing software should be commercially available. No WIP solution is accepted. | Deleted |

Modifications in Technical Specifications has been uploaded on the website
<https://eprocure.gov.in/eprocure/app>.

Further please note that at this juncture all other Terms & Conditions of the tender remains same as per the Tender Documents.

Purchase Officer