

TECHNICAL COMPLIANCE FORM			
Ref: Tender No:			
TATA MEMORIAL CENTRE			
ADVANCED CENTRE FOR TREATMENT, RESEARCH & EDUCATION IN CANCER			
TECHNICAL SPECIFICATION OF :- - " Vortexer Mixer With Accessories- Qty - 3 Nos."			
Intended Use: For Radio pharmacy lab in Nuclear Medicine Dept.			
Sr. No.	Name of the Vendor :		
A	Technical Specifications / Scope of supply	COMPLIANCE 1) Mention "Complied /Not Complied" 2) Highlight any deviations. 3) Mention part number/Catalogue number of the relevant item quoted.	REMARKS
B	Name of the Manufacturer :		
C	Country of Origin:		
D	Model Name / Model No. of the Equipment (If the equipment is a combination of more than one equipment then list down the model no. pertaining to each):		
E	Year of Introduction Internationally		
F	Year of Introduction in India		
G	Please quote the latest state of the art model which meets the tender specifications. The offer should be complete in all respect for the intended use.		
H	Technical Specifications/Scope of supply		
I	" Vortexer Mixer with Accessories" , complete as per below mentioned configuration and specifications:		
J	Features and specifications required:		
1	RPM Range :- 0 to 3,000 rpm		
2	Attachments for vial, eppendorf tubes, conical flask, plates etc		
3	Vortex shaker with speed control		
4	Spare Cup Attachment, Spare One hand attachment, Spare One hand insert, Spare Micro tube insert.		
6	Standard accessories (Which comes along with the machine, pricing of which is part of the equipment)		
7	Periodic Calibration		
8	Safety requirements:It should follow International / national safety requirement. Please specify certification with certifying agency and country with copies of certificate.		

9	The Equipment/ System should comply to US FDA/ CE/ BIS/ ICMED certification. Please attach copy of the certificate(s).		
10	User's list: A list of installations with the address and contact numbers to be provided. (User list should be for the quoted model.)		
11	Input power supply requirements: 240 V AC \pm 10 %, 50 Hz, single phase. Specify if any other power supply requirement is recommended.		
12	Warranty and after sales support		
13	Warranty : 2 Years comprehensive-on site- factory warranty for the entire equipment system. QA/QC as per prescribed protocol during the warranty period.		
14	Important terms to be noted by the bidders:		
15	Manufacturer/bidder should clearly indicate typical life span / Life expectancy of the quoted equipment.	-- years	
16	Mention typical site requirements for the installation.	--	
17	Declaration by the bidder		
18	We have quoted for all the items meeting the description/scope of supply in the Financial bid as per prescribed format of the Tender documents and we agree that Partial/incomplete offers are liable for rejection.		