

Tata Memorial Centre			
Advanced Centre for Treatment, Research and Education in Cancer			
Technical Specifications Of: Vertical Autoclave - 1 No			
Sr No	Scope of supply	Quantity	COMPLIANCE 1) Compliance column may be filled with relevant data, figures, range etc. as applicable. Do not just mention "YES / NO / Complied". 2) Highlight any deviations. 3) Mention part number/Catalogue number of the relevant item quoted.
Technical Specifications / Scope of Supply			
1	Vertical Autoclave, complete as per below mentioned configuration and specifications:B21B7:C2B7:C20	1 No.	
2	Size/ Chamber dimension - 18 inch diameter, 24 inch depth		
3	Working pressure- 1.2 kgf/cm sq.g (15 psi - 17 psi)		
4	Internal chamber Capacity :- 50 Litre.		
5	Sterilizing. Temperature- 121°C to 134°C Heat Average.		
6	Construction- Triple wall & SS GMP construction. (GMP Certification should be attached)		
7	Outer chamber- 304 stainless Steel		
8	Inner chamber- 304 SS		
9	Insulation- Glass Wool		
10	Door locking- Radial Type -Wing Nuts		
11	Silicon gasket.		
12	Power supply- 220 Volts / 440 Volts		
13	Load- 3 KVA		
14	Fitted with safety valves, steam release valve, pressure hauges, Temperature gauges, safety thermostat for over temperature., water filling port, water level indicator, etc.		
15	SS mesh basket to be supplied.		
16	Hydraulic testing certificate to be provided.		
17	NOTE: Technical literature/Brochure/catalogue of the quoted model shouyld be attached with the bid.		
18	Supply, installation, commissioning and testing is to be included in the scope of supply.		
19	Warranty: 2 years comprehensive-on site- Factory warranty		Agree/Not Agree
20	Important terms to be noted by the bidders:		
21	<u>Read the above scope of supply carefully and quote accordingly. Incomplete and /or partially complete offers are liable to be rejected.</u>		Agree/Not Agree
22	<u>After Sales Service Support:</u> Bidder must have authorized service centre in Mumbai/Navi Mumbai/Thane Areaa. Provide address and staff strength.		Yes/No
23	Item/Equipment quoted should not be a second hand/refurbished/on verge of obsolescence/outdated model. Similarly, after award of contract, the item supplied should be a newly manufactured item and not the old item already manufactured and stocked for more than 1 month.		Agree/Not Agree
24	After opening of the Technical bid (Part-1), Physical demonstration of the quoted model may have to be shown / arranged by the bidder, if requested by the Institute. Physical demonstration may be shown at ACTREC within 10 days from the date of request. Physical Demonstration may be requested to confirm the availability of any or all technical features as mentioned/stated in the technical bid. Physical Demonstration will also be a part of technical evaluation process. If the bidder does not comply, such bids are liable to be disqualified. (Demonstration of quoted model is to be shown and not the demonstration of similar models with different technical specifications and features)		Agree/Not Agree

25	Past experience of the bidders in terms of quality of supplied equipments, after sales service and application support will be taken into consideration while technical evaluation.. Bidders who has unsatisfactory past experience in last 2-3 years, in terms of quality of supplied equipments, after sales service and application support, bids of such bidders may liable to be rejected.		Agree/Not Agree
26	Complete and detailed information should be provided in respect to each point specified in the specifications. <u>Technical bids that are incomplete in any respect are liable to be rejected.</u> Provide relevant supportive information, publications, catalogue, etc. Bidders providing misleading or wrong information are liable to be rejected. All technical claims should be printed in the technical brochure of the equipment.		Agree/Not Agree
27	If any contradictory statements /figures/information is observed in the compliance chart and in the technical bid, then the technical information mentioned in the product literature/brochure will be considered true and further evaluation will be done based on the information given in the product literature/brochure. Regarding the warranty and free AMC terms, information quoted in the technical bid will only be taken as valid terms.		Agree/Not Agree
28	Compliance column may be filled with relevant data, figures, range etc. as applicable. Do not just mention "YES / NO / Complied ".		Agree/Not Agree
29	Provide list of current installations(in the last 2-3 years) of the quoted model in India. Contact details of the end users should be provided. Provide name of the end user, email ID, tel. no. etc.		Agree/Not Agree
30	Declaration by the bidder		
31	We have quoted for all the items meeting the description/scope of supply in the Financial bid as per prescribed format of the Tender documents and we agree that Partial/incomplete offers are liable for rejection.		Yes/No