

TECHNICAL COMPLIANCE FORM			
Ref: Tender No:			
TATA MEMORIAL CENTRE			
ADVANCED CENTRE FOR TREATMENT, RESEARCH & EDUCATION IN CANCER			
TECHNICAL SPECIFICATION OF :- "Suction Pump ( Vacuum Pump ) - 2 Nos.			
SR. NO.	Name of the Vendor :		
A)	Technical Specifications / Scope of supply	Quantity	<b>COMPLIANCE</b> 1) Mention "Complied /Not Complied" 2) Highlight any deviations. 3) Mention part number/Catalogue number of the relevant item quoted.
B)	Name of the Manufacturer :		
C)	Model No. of the Equipment		
D)	Year of Introduction Internationally		
E)	Year of Introduction in India		
F)	Technical Specifications/Scope of supply		
1	Suction pump ( Vacuum Pump) , complete as per below mentioned configuration and specifications:	2 Nos.	
2	Compact, reliable, lightweight, oil free, diaphragm pump with low noise level and vibration free operation.		
3	Suction feet / rubber cups for wobble free operation and reducing vibrations.		
4	Suitable for applications like filtration, suction, etc.		
5	Max achievable vacuum : approx. 600 to 630 mm Hg.		
6	Max Flow Rate: approx. 20 Ltr/Min		
7	Pump HP: Approx. 1/6 HP or 1/8 HP		
8	Pump should be fitted with "barbed nozzle" at the outlet side.		
9	Vacuum dial at outlet side.		
10	Documentation		
11	Provide technical literature/Brochure of the quoted model and picture of the same.		
12	Input power supply requirements: 240 V AC $\pm$ 10 %, 50 Hz, single phase . Suitable power adaptor/ eliminator should be supplied, if necessary.		
13	Warranty: 2 year comprehensive-on site- Factory warranty- for entire equipment.		
14	Important terms to be noted by the bidders:		
15	<u>Read the above scope of supply carefully and quote accordingly. Incomplete and /or partially complete offers are liable to be rejected.</u>		
16	Bidders should clearly mention about compliance of the warranty and AMC/free service terms in the technical bid. Bids not fully complying to the warranty and AMC/free service terms are liable to be rejected.		
17	Item/Equipment quoted should not be a second hand/refurbished/on verge of obsolescence/outdated model. Similarly, after award of contract, the item supplied should be a newly manufactured item and not the old item already manufactured and stocked for more than 1 month.		
18	After opening of the Technical bid (Part-1), Physical demonstration of the quoted model may have to be shown / arranged by the bidder, if requested by the Institute. Physical demonstration may be shown at one of the end user's site/Principle company's application lab/manufacturing site located in Mumbai/Navi Mumbai/Thane cities. If there are no installations of the quoted model in Mumbai/Navi Mumbai/Thane cities, then the quoted model may have to be brought in at ACTREC for demo purpose within 10 days from the date of request. Physical Demonstration may be requested to confirm the availability of any or all technical features as mentioned/stated in the technical bid. Physical Demonstration will also be a part of technical evaluation process. If the bidder does not comply, such bids are liable to be disqualified. (Demonstration of quoted model is to be shown and not the demonstration of similar models with different technical specifications and features)		

19	Past experience of the bidders in terms of quality of supplied equipments, after sales service and application support will be taken into consideration while technical evaluation.. Bidders who has unsatisfactory past experience in last 2-3 years, in terms of quality of supplied equipments, after sales service and application support, bids of such bidders may liable to be rejected.		
20	Complete and detailed information should be provided in respect to each point specified in the specifications. <u>Technical bids that are incomplete in any respect are liable to be rejected.</u> Provide relevant supportive information, publications, catalogue, etc. Bidders providing misleading or wrong information are liable to be rejected. All technical claims should be printed in the technical brochure of the equipment.		
21	If any contradictory statements /figures/information is observed in the compliance chart and in the technical bid, then the technical information mentioned in the product literature/brochure will be considered true and further evaluation will be done based on the information given in the product literature/brochure. Regarding the warranty and free AMC terms, information quoted in the technical bid will only be taken as valid terms.		
22	Compliance column may be filled with relevant data, figures, range etc. as applicable. Do not just mention "YES / NO / Complied".		
23	Provide list of current installations( in the last 2-3 years) of the quoted model in India. Contact details of the end users should be provided. Provide name of the end user, email ID, tel. no. etc.		
24	Declaration by the bidder		
25	We have quoted for all the items meeting the description/scope of supply in the Financial bid as per prescribed format of the Tender documents and we agree that Partial/incomplete offers are liable for rejection.		