TECHNICAL COMPLIANCE FORM Ref: Tender No: ACTREC/PUR/2022-2023/TC/185 **TATA MEMORIAL CENTRE** ADVANCED CENTRE FOR TREATMENT, RESEARCH & EDUCATION IN CANCER TECHNICAL SPECIFICATION OF :- - " Slide Warmer- 4 Nos. " Intended Use: For heating the Tissue slides in Histopathology lab Sr. No. Name of the Vendor: Α Technical Specifications / Scope of supply COMPLIANCE Remarks (to provide 1) Mention "Complied /Not details) Complied" 2) Highlight any deviations. 3) Mention part number/Catalogue number of the relevant item quoted. Name of the Manufacturer: В C Country of Origin: D Model Name / Model No. of the Equipment (If the equipment is a combination of more than one equipment then list down the model no. pertaining to each): Ε Year of Introduction Internationally Year of Introduction in India F Please quote the latest state of the art model which meets the G tender specifications. The offer should be complete in all respect for the intended use. Н **Technical Specifications/Scope of supply** 1 " Slide Warmer-1 No. ", complete as per below mentioned configuration and specifications: Features and specifications required: J LCD display should be available. Feather touch keypad for setting plate temperature 3 Temperature range should be: From Ambient to 90 degC 4 Temperature tolerance should be +/-1 degC Dimension of the plate should be approximately 350 X 250mm Over all dimension should not exceed 360 X 320X 120 (LXBXH) 6 Power consumption should be less than 250W. 8 Performance parameter required as below: (like accuracy, sensitivity, sample volume requirement, resolution or such performance parameters to be highlighted) IQ/OQ/PQ should be provided Periodic Calibration 10 11 Regulatory Approvals, If any. Details with copies of approvals. The Equipment/ System should comply to US FDA/ European CE/ BIS/ ICMED certification. Please attach copy of the certificate(s). Input power supply requirements: 240 V AC ± 10 %, 50 Hz, single 13 phase. Specify if any other power supply requirement is recommended. 14 Warranty and after sales support Warranty: 2 Years comprehensive-on site-factory warranty for the entire 15 equipment system. Manufacturer/bidder should clearly indicate typical life span / Life -- years expectancy of the quoted equipment. Mention typical site requirements for the installation. 17 Item/Equipment quoted should not be a second hand/refurbished/on Agree/Not Agree verge of obsolescence/outdated model. Similarly, after award of contract, the item supplied should be a newly manufactured item and not the old item already manufactured and stocked for more than 1

month.

19	After opening of the Technical bid (Part-1), Physical demonstration of the quoted model may have to be shown / arranged by the bidder, if requested by the Institute. Physical demonstration may be shown at one of the end user's site/Principle company's application lab/manufacturing site located in Mumbai/Navi Mumbai/Thane cities.	Agree/Not Agree	
	If there are no installations of the quoted model in Mumbai/Navi Mumbai/Thane cities, then the quoted model may have to be brought in at ACTREC for demo purpose within 10 days from the date of request. Physical Demonstration may be requested to confirm the availability of any or all technical features as mentioned/stated in the technical bid. Physical Demonstration will also be a part of technical evaluation process. If the bidder does not comply, such bids are liable to be disqualified. (Demonstration of quoted model is to be shown		
	and not the demonstration of similar models with different technical specifications and features)		
20	Compliance column may be filled with relevant data, figures, range etc. as applicable. Do not just mention "YES / NO / Complied ".	Agree/Not Agree	
21	Declaration by the bidder		
22	We have quoted for all the items meeting the description/scope of supply in the Financial bid as per prescribed format of the Tender documents and we agree that Partial/incomplete offers are liable for rejection.	Yes/No	